

SPECTRUM TECHNIQUES Inc.

106 Union Valley Road
Oak Ridge, TN 37830 USA

Phone: 865-482-9948

Fax: 865-483-0473

E-mail: sales@spectrumtechniques.com

Web site: www.spectrumtechniques.com

Certificate of Compliance

This is to certify that the enclosed Check Source complies fully with USNRC Regulations 10CFR part 30.18 Exempt Quantities, does not exceed the activity as set forth in section 30.71 Schedule B and may be transferred to any individual free of specific licensing requirements. The package conforms to USDOT 49CFR 173.421, Postal Publication No. 6 (dated September 6, 1983) and IATA section 4.2 Identification (UN2910) Radioactive material, excepted package, limited quantity of material, N.O.S. (not otherwise specified), Class 7, section 10.5.9.4 calculating A2 value of radionuclide for packaging and transportation.

This is to certify that the enclosed Check Source was manufactured:

August, 2010

Sample

At the time of manufacture, the source contained a nominal activity of 1.0 microcuries of Cs-137 which decays with a half-life of 30.2 years.

The source is of sealed construction and wipe tested for surface contamination and leakage.

Larry Webb

Larry Webb, Radioisotope Manager.

August, 2010

Dated: